Near-Infrared High-Throughput Spectrograph (NIHTS)

Instruments

Facility

Instrumentation M ain Page

Quick Links

LDT Observer Information Home

Instrumentation Applying for

Observing Time

Observing Run Planning

At the Telescope (Printable Logsheets)

Target of **Opportunity (ToO)** Policies

After Your **Observing Run**

First-Time Users

Logistics

LDT Important Notes

Instruments: Large Monolithic Imager (L **DeVeny** Facility Spectrog LDT Science NIR Schedule Spectrog (NIHTS) LDT Staff Telescope Visitor / Instrume Site Information **High-Res** Weather Spectrog (EXPRES Acknowledging I DT Speckle I (QWSSI) Selected Technical **High Spe** Publications Imager (F NIR Spectrog (RIMAS)

NIHTS internal layout (NIHTS Manual, fig. 25).

_MI)		
Optical raph	/ Observing Planning	At the Telescope
raph	Applying for Observing Time	Observing Quick Links
DI	- Planning Checklist	Target of Opportunity (ToO Policies)
ents:	Remote Observing	Remote Observing
; raph S)	Opportunity (ToO) Policies	LDT Startup Procedures
mager	Target Lists	Closure Conditions
	First-Time Users	Night Feedback
ed POETS)	Logistics	End of Night Tasks
raph		

User Manuals:

LMI Manual **DeVeny Manual**

NIHTS Manual

LDT Facility Instrument – Available Full-Time, Remote Available

Mounting Port: Instrument Cube - Port D (small)

Manual: NIHTS User Manual v1.7 (26 April 2021)

Additional Information: Overview of NIHTS (2018 Proc. SPIE 10702, 107023E), NIHTS Commissioning Paper (Gustafsson, A. et al. 2021, PASP, 133, 035001)

Instrument Scientist: Nick Moskovitz (nmosko at lowell dot edu)

The Near-Infrared High-Throughput Spectrograph (NIHTS, pronounced "nights") is a low-resolution (R~200) near-infrared prism spectrograph, covering 0.86-2.4 microns in a single setup at a resolution of ~100 to 200. NIHTS contains no moving parts and employs a single slit mask that offers 7 different width slits (4.03, 1.34, 0.81, 0.27, 0.54, 1.07 and 1.61 arcseconds), each approximately 10 arcseconds in length. The instrument is fed by a dichroic fold flat on a linear translation stage in the instrument cube, which allows simultaneous LMI imaging and NIHTS spectroscopy.

Users scheduled with NIHTS are encouraged to contact Nick Moskovitz at least two weeks ahead of their run to provide information about the planned observations. Afternoon and night time support for the instrument will be available for first-time users. A data reduction pipeline for NIHTS developed by M. Cushing (U. Toledo) and based on the IDL Spextool package is available.



NB: (2024Mar25) LDT staff are in the process of adding NIHTS to the Pypelt spectroscopic data reduction pipeline. This will be an alternate data reduction pathway to the existing SpeXTool pipeline in IDL, and will produce reduced data in the same format as for DeVeny.

Instrument Quick Facts:

Detector: Rockwell Hawaii-1 array: 1024 x 1024 pixels

Slit Length: 12" (per slit segment)

Pixel Scale: 0.13"/pixel (spatial direction)

Gain: 24 e⁻/ADU

Read Noise: 100 e-

Normal Operating Temperature: 75 K

Readout Time: 0.001504 sec (Row Time)

Minimum Exposure Time: 0.77 sec

Slit Widths: 4.03", 1.61", 1.34", 1.07", 0.81", 0.54", 0.26"

Slit Viewing Camera:

Model: Xenics XEVA IR camera

Pixel Scale: 0.326"/pixel

Image Size: 320 × 256 pixels

Field of View: 60" × 90"